

CLIENT LOGO

COMPANY NAME

PLANT NAME

QA Department



Scanner\_Data\_Log

Department	DEPARTMENT NAME			
Filter From	12-Dec-2024 12:00	Filter To	12-Dec-2024 12:20	
Records From	12-Dec-2024 12:00	Records To	12-Dec-2024 12:20	
Equipment	Photo Stability 04CH		Code	001
Temperature °C	Low Alarm	0.0	High Alarm	28.0
Humidity %RH	Low Alarm	0.0	High Alarm	55.5

Channel Name	CH	Minimum	Maximum	Average
Temperature1	T_CH1	25.60	26.50	26.08
Temperature2	T_CH2	25.60	26.50	26.09
Temperature3	T_CH3	25.60	26.50	26.00
Temperature4	T_CH4	25.50	26.50	25.99

Channel Name	CH	Minimum	Maximum	Average
Humidity1	H_CH1	50.00	50.90	50.48
Humidity2	H_CH2	50.00	50.90	50.40
Humidity3	H_CH3	50.10	50.90	50.56
Humidity4	H_CH4	50.10	51.00	50.52

CLIENT LOGO

COMPANY NAME

PLANT NAME

QA Department



Scanner\_Data\_Log

Department	DEPARTMENT NAME			
Filter From	12-Dec-2024 12:00	Filter To	12-Dec-2024 12:20	
Records From	12-Dec-2024 12:00	Records To	12-Dec-2024 12:20	
Equipment	Photo Stability 04CH		Code	001
Temperature °C	Low Alarm	0.0	High Alarm	28.0
Humidity %RH	Low Alarm	0.0	High Alarm	55.5

Sr. No.	Log Date	Time	T_CH1	T_CH2	T_CH3	T_CH4	H_CH1	H_CH2	H_CH3	H_CH4
1	12-Dec-2024	12:00	26.3	25.6	26.5	25.8	50.5	50.9	50.7	50.1
2	12-Dec-2024	12:01	25.8	26.1	26.5	26.3	50.0	50.4	50.7	50.6
3	12-Dec-2024	12:02	25.8	25.6	25.6	26.0	50.7	50.0	50.9	50.2
4	12-Dec-2024	12:03	25.9	26.3	26.1	26.5	50.2	50.5	50.9	50.7
5	12-Dec-2024	12:04	26.4	25.8	26.1	26.0	50.2	50.0	50.4	50.2
6	12-Dec-2024	12:05	26.4	26.3	25.6	26.0	50.7	50.5	50.9	50.2
7	12-Dec-2024	12:06	25.9	26.3	26.1	26.5	50.2	50.6	50.4	50.7
8	12-Dec-2024	12:07	26.4	25.8	25.6	26.0	50.7	50.1	50.4	50.2
9	12-Dec-2024	12:08	25.9	26.3	25.7	26.5	50.7	50.5	50.9	50.7
10	12-Dec-2024	12:09	26.0	25.8	26.2	25.5	50.2	50.0	50.4	50.8
11	12-Dec-2024	12:10	26.5	25.8	25.7	26.0	50.7	50.1	50.9	50.3
12	12-Dec-2024	12:11	26.0	26.3	26.2	25.5	50.2	50.6	50.9	50.8
13	12-Dec-2024	12:12	26.5	25.8	26.2	26.0	50.2	50.1	50.4	50.3
14	12-Dec-2024	12:13	26.5	26.3	25.7	26.0	50.7	50.6	50.9	50.3
15	12-Dec-2024	12:14	26.0	26.3	26.2	25.5	50.2	50.6	50.4	50.8
16	12-Dec-2024	12:15	25.6	26.5	25.8	25.7	50.9	50.7	50.1	50.4
17	12-Dec-2024	12:16	26.1	26.0	26.3	25.7	50.4	50.7	50.6	50.9
18	12-Dec-2024	12:17	25.6	26.0	25.8	26.2	50.9	50.2	50.1	50.4
19	12-Dec-2024	12:18	26.1	26.5	25.9	25.7	50.4	50.7	50.1	50.9

CLIENT LOGO

COMPANY NAME

PLANT NAME

QA Department



Scanner\_Data\_Log

Department	DEPARTMENT NAME			
Filter From	12-Dec-2024 12:00	Filter To	12-Dec-2024 12:20	
Records From	12-Dec-2024 12:00	Records To	12-Dec-2024 12:20	
Equipment	Photo Stability 04CH		Code	001
Temperature °C	Low Alarm	0.0	High Alarm	28.0
Humidity %RH	Low Alarm	0.0	High Alarm	55.5

Sr. No.	Log Date	Time	T_CH1	T_CH2	T_CH3	T_CH4	H_CH1	H_CH2	H_CH3	H_CH4
20	12-Dec-2024	12:19	26.2	26.0	26.3	26.2	50.4	50.2	50.6	51.0
21	12-Dec-2024	12:20	25.7	26.5	25.8	26.2	50.9	50.3	50.1	50.5